

## Report Details

Report Location C:\Users\Didattica\Desktop\chim\_fis\_2\2016\c5c7\polistirene\_1\_2.pdf  
Report Creator user  
Report Date Thursday, June 16, 2016 10:06 AM

## Sample Details

Sample Name 1  
Sample Description Sample 019 By ir Date Thursday, June 16 2016  
Analyst user  
Creation Date 6/16/2016 10:00:18 AM  
X-Axis Units cm-1  
Y-Axis Units A

## Instrument Details

Instrument Model Spectrum Two  
Instrument Serial Number 100169  
Software Revision NIOS2 Main 00.02.0079 15-December-2014 11:35:32  
Number of Scans 4  
Resolution 1

## Instrument Details (Full)

Instrument Model Spectrum Two  
Instrument Serial Number 100169  
Software Revision NIOS2 Main 00.02.0079 15-December-2014 11:35:32  
Number of Scans 4  
Resolution 1  
Detector LiTa03  
Source MIR  
Beamsplitter OptKBr  
Apodization Strong  
Spectrum Type Spectrum  
Beam Type Ratio  
Phase correction Self  
Scan Speed 0.2  
IGram Type Single  
Scan Direction Combined  
Zero Crossings 0  
JStop 4.47  
IR-Laser Wavenumber 11750.00  
Manufacturer L1600217  
Part Number L1600217  
Description Sample base plate assy (non RFID)  
Default Scan Range / cm-1 4000 450  
Temperature / °C Not Specified  
Accessory Type Slide Holder  
Slide Holder Option Not Specified

## Accessory

Manufacturer L1600217  
Part Number L1600217  
Description Sample base plate assy (non RFID)  
Default Scan Range / cm-1 4000 450  
Temperature / °C Not Specified  
Accessory Type Slide Holder  
Slide Holder Option Not Specified


## Quality Checks

Water Vapor	Passed
Carbon Dioxide	Passed
Baseline Low	Passed
Baseline High	Passed
Baseline Slope	Passed
Strong Bands	Warning
Weak Bands	Passed
High Noise	Passed
Fringes	Warning
Vignetting	Passed
Blocked Beam	Passed
Negative Bands	Passed
Zero Transmission	Passed
Stray Light	Passed
Window Cutoff	Passed

### History

Who	What	When	Parameters	Comment
user	Created as New Dataset	6/16/2016 10:00:18 AM		Sample 019 By ir Date Thursday, June 16 2016
user	Atmospheric Correction	6/16/2016 10:00:18 AM		
user	Absorbance	6/16/2016 10:00:18 AM	"Channel:1", "Result.sp"	

### Spectrum



Name	Description
1	Sample 019 By ir Date Thursday, June 16 2016

### Summary

Sample Name	Description	Quality
1	Sample 019 By ir Date Thursday, June 16 2016	The Quality Checks give rise to multiple warnings for the sample.